INSPECTION CERTIFICATE
TRACEABLE TO NIST

LARGE DIAL FACE DIAL INDICATORS

2312-15



- Meet ASME B89.1.10M-2001
- With lug back
- Supplied with limit pointers and bezel clamp
- Optional accessory: contact points (page 73)

Part No.	Range	Graduation	Accuracy		Range/rev	Dial reading
			First 2.5 rev	Overall		
2312-15 *	1.5"	.001"	±.001"	±.004"	.1"	0-100
2312-2	2"	.001"	±.001"	±.004"	.1"	0-100

*Supplied with manufacturer inspection certificate traceable to NIST