



www.insizeus.com

TOLL FREE: 844-520-7668

ISO17025 CALIBRATION SERVICES

INSIZE Calibration Services  
1200 Woodruff Road, Suite A-16  
Greenville, SC 29607  
Phone: 864-284-6262  
Fax: 864-284-6212

Page 1 Of 1

**Certificate Of Calibration** Cert. #: 6780

Calibration Performed By: INSIZE CALIBRATION SERVICES  
1200 WOODRUFF ROAD, SUITE A-16  
GREENVILLE, SC 29607

For: TEST COMPANY  
1234 MAIN ST  
ANYTOWN, USA

Serial Number: 12345  
Gage Description: .003" DIAL INDICATOR .0001"/DIV  
Gage Type: INDICATOR  
Procedure Used: SCP-09 INDICATORS  
Calibration Date: 03/10/2016  
Calibration Result: Passed

Gage I.D.: 12315  
Unit of Measure: INCH  
Performed By: cringage  
Calibration Due Date: 03/10/2017

Seq.	Reference Item	Standard	Tol. +	Tol. -	As Found	Final	Unit
1	ASCENDING 25%	0.00200	0.00212	0.00190	0.00209	0.00209	INCH
2	ASCENDING 50%	0.00400	0.00413	0.00390	0.00400	0.00400	INCH
3	ASCENDING 75%	0.00600	0.00613	0.00590	0.00600	0.00600	INCH
4	ASCENDING 100%	0.00800	0.00813	0.00790	0.00800	0.00800	INCH
5	REPEAT ZERO	0.00000	0.00013	-0.00019	0.00000	0.00000	INCH
6	DESCENDING 100%	0.00600	0.00613	0.00590	0.00600	0.00600	INCH
7	DESCENDING 75%	0.00400	0.00413	0.00390	0.00400	0.00400	INCH
8	DESCENDING 50%	0.00200	0.00213	0.00190	0.00200	0.00200	INCH
9	REPEAT ZERO	0.00000	0.00013	-0.00019	0.00000	0.00000	INCH

F = Out Of Tolerance L=Out of Tolerance/Limits Gage Use R=Report Of Value A=Adjusted

Equipment Used To Calibrate Gage:  
This calibration is performed in a controlled environment with guidelines compliant to ISO/IEC 17025:2005 and ANSI Z-540-1:1994. This calibration was conducted using standards traceable to the BIP through NIST.  
In tolerance conditions are based on the calibration falling within the specified limits and are not affected by the measurement uncertainty. Actual uncertainty represents the expanded uncertainty at approximately 95% confidence level using a k2 factor. Any number of conditions may cause this calibration to be invalidated prior to the established due cycle.  
This report shall not be reproduced, except in full, without the written approval of INSIZE Calibration Services.

Approved By:  
*Lee Cook*  
Signature:  
Lee Cook  
Calibration Metrologist

- INSIZE is proud to announce our new Calibration Service "INSIZE Calibration Services"
- The most popular tools ready to ship same day!
- Tools ordered by noon will ship same day with an ISO17025 Certificate with your company name, address, and date of service. No need to wait a week or more
- All INSIZE Precision Tools available with ISO17025 certification. Contact us for pricing on all products

Tools Available to Ship Today with Calibration Certificate

Part No.	Description	List Price
1108-150CAL	0-6"/0-150mm Electronic Caliper	\$89.00
1108-200CAL	0-8"/0-200mm Electronic Caliper	\$115.00
1108-300CAL	0-12"/0-300mm Electronic Caliper	\$159.90
1311-6CAL	0-6" Dial Caliper	\$88.50
1311-8CAL	0-8" Dial Caliper	\$106.00
1311-12CAL	0-12" Dial Caliper	\$159.50
3203-1ACAL	0-1" Outside Micrometer	\$69.70
3203-2ACAL	1-2" Outside Micrometer	\$88.70
2307-1CAL	0-1" Dial Indicator	\$69.00

Distributed by:



**INSIZE USA**  
Add: 3492 Diversified Drive  
Loganville, GA 30052  
Tel: 470-240-2520  
Fax: 470-448-4771  
E-mail: info@insizeus.com